

01	Ensure patient is a qualified candidate for Zerona, discuss package options, sign accountability contract [if applicable]	
02	Body composition analysis [if applicable], pre- measurements, and before pictures have been completed	
03	Make sure treatment area is exposed [treatment should be on bare skin]	
04	Place center diodes over the area of greatest fat accumulation	
05	Ensure laser diodes are 3-6 inches away from the surface of the skin [use tape measure to spot check]	
06	Take measures to ensure only areas wanting to be treated are exposed to laser. Gown or drape appropiately and adjust laser positioning as needed	
07	When laser pauses at halfway point have patient flip over, re-position laser diodes and spot check distance	
08	Upon patient completing treatment ensure they understand aftercare recommendations	
09	Give patient checklist handout with aftercare recommendations	
10	Progress measurements taken mid package, post	

measurements taken and after pictures taken